


VERIFY STEAM Integrating Indicators:

The VERIFY STEAM Integrating Indicator is used on its own or in conjunction with a biological indicator for routine monitoring and release of steam sterilized loads.



During the test cycle, the steam sensitive chemical located within the strip will wick across the “REJECT” viewing window and enter into the “ACCEPT” viewing window. Sufficient steam has entered the reusable test pack when any of the dark bar is visible within the “ACCEPT” Viewing window.

